

Description: Disposable latex examination gloves, ambidextrous, beaded cuff, non-sterile, powder-free, Micro-Textured.

Standards: ASTM D 3578-95
 ASTM D 5151-92 Detection of holes in medical gloves
 ASTM D 6124-97
 EN 455
 EN 420
 Federal Register Part 820 (GMP), Pre-Market Notification 510k
 FDA Compliance Policy Guide No. 7124.31, Chapter 24 Devices.
 FDA 1,000ml water tight test rejection at S-4: 1.5 AQL maximum.
 Title 21 CFR 177.26 for Food Handling
 ISO 9002

Powder Free Process: Chlorination

Other certifications, standards & memberships:



Storage: Avoid direct sunlight. Store in a cool and dry place.

Tensile Strength (Mpa):	Before aging	Minimum 17
	After aging at 70°C, 7 days	Minimum 16
Ultimate Elongation %:	Before aging	Minimum 750
	After aging at 70°C, 7 days	Minimum 750
Modulus of Elasticity (Mpa)	Before aging	Minimum 1.3
At 300% elongation	After aging at 70°C, 7 days	Minimum 1.5
Cuff Strength	Before aging	Minimum 300%

Size	X-Small	Small	Medium	Large	X-Large	XX-Large
Width (± 2 mm) min.	73	84	94	103	114	n/a
Length (± 3 mm)	245	245	245	245	245	n/a
Thickness						
Cuff (±0.02 mm)	0.12	0.12	0.12	0.12	0.12	n/a
Palm (±0.02 mm)	0.15	0.15	0.15	0.15	0.15	n/a
Finger (±0.02 mm)	0.18	0.18	0.18	0.18	0.18	n/a
MIL (0.001")	5	5	5	5	5	n/a

CAUTION / WARNING

- Components used in making gloves may cause allergic reactions in some users.
- In storage avoid excessive heat (40°C, 104° F). Open box should be shielded from exposure to direct sun or fluorescent lighting.
- Good quality latex gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.
- This product contains natural rubber latex which may cause allergic reactions.
- Please seek medical advice before using these gloves if you or your patient suffer from or suspect allergic reactions. Prolonged usage of latex gloves can result in sensitization to latex proteins.
- These gloves are sometimes subject to failure and may puncture or tear if donned improperly by being in contact with sharp fingernails or objects such as finger rings, etc.
- The buyer and user have examined these gloves after having read the cautionary statements above, and have determined them to be fit for the purpose intended. These gloves comply with US ASTM, US FDA and CE standards, and are intended for general purpose medical use in non-sterile environments. The manufacturer, its resellers, agents and their employees are not responsible for any other claims or representations, in addition to, or to the contrary.

MATERIAL SAFETY DATA SHEET (MSDS)

(This section serves as MSDS)

HEALTH HAZARD DATA

The chemical formulation of the gloves and the chlorination process do not contain any substances known to be harmful to the user. Never the less, you are required to follow any cautionary and warning labels on the product packaging.

Some users may be allergic to components used in the manufacturing process. If this is the case, you must immediately stop using this product and consult your medical doctor.

These gloves have undergone BIOCOMPATABILITY tests as recommended by the US FDA, for Primary Skin Irritation and Dermal Sensitization, and were found to be non-irritant and non-sensitive.

This product is FREE FROM NATURAL RUBBER LATEX PROTEINS.

474 Main Avenue
 PO Box 83
 Neche, ND 58265